

AUG 13 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130142</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>SCOTT GARDNER</u>	ID# <u>6222</u> ✓ Date <u>3 Aug 10</u> ✓
A Agency <u>FT Richardson (FRN 4)</u>	Phone # <u>384-3888</u>
Instrument Location <u>Bldg 656, A St FT Richardson</u>	
Alco S/N <u>31085</u> ✓ Target Value <u>.074</u> ✓ High Pressure <u>1000 psi</u>	
B Alco Test Values	<u>.072</u> ✓ 1st Alco
	<u>.076</u> ✓ 2nd Alco
Signature <u>Scott Gardner</u>	<u>CAS</u> <u>8/14/10</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of SEPT, 2010.

Daniel L. Nicholas  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N

130142  
CAB 8/16/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130142

AUGUST 03, 2010

CAB 8/16/10

OPERATOR'S NAME:

GARDNER/SCOTT/E

OPERATOR'S NUMBER: 6222

SUBJECT'S LAST NAME:

N/A

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: N/A

DEPT/AGENCY: FRM

CASE/REPORT: VERIFICATION OF CAL

TEST TYPE: U

ALCO TARGET VALUE: .074

ALCO S/N: 31085

### — BREATH ANALYSIS —

.074 ADJUSTED FOR 30.19 in		
ALCO TARGET	.074	14:51
BLANK TEST	.000	14:52
INTERNAL STANDARD	VERIFIED	14:52
ALCO TV 30.17 in	.072	14:52
BLANK TEST	.000	14:53
SUBJECT SAMPLE	.000	14:54
BLANK TEST	.000	14:54
ALCO TV 30.19 in	.076	14:55
BLANK TEST	.000	14:56

CAB 8/16/10

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY

DATAMASTER cdm 130142

AUGUST 03, 2010

TIME 14:47

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	44c
BAROMETER:	30.17 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

'"#\$%&'()\*+,-./0123456789:;<=?@ABCDEFGHI  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
prstuvwxyz{|}~